

Certificate of Compliance

Certificate Number 240605 - E238404
Report Reference E238404, June 6th, 2005
Issue Date 2005 June 24

Page 1 of 1



Issued to: **LMT LAMMERS MEDICAL TECHNOLOGY GMBH**
OSTERWEIDE 8
D-23562 LUEBECK GERMANY


This is to certify that representative samples of: **Transport Incubator**
LMT model nomag IC 1.5


Have been investigated by Underwriters Laboratories Inc.® in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: **UL 60601-1 - Medical Electrical Equipment, Part 1: General Requirements for Safety**
CSA C22.2 No.601.1- Medical Electrical Equipment, Part 1: General Requirements

Additional Information: **Electrical Ratings:**
120Vac, 50/60Hz; 230Vac, 50/60 Hz, 475W

Only those products bearing the UL Classification Mark for the U.S. and Canada should be considered as being covered by UL's Classification and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Classification Mark includes: the UL in a circle symbol:  with the word "CLASSIFIED" (as shown), a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) as indicated in the appropriate UL Directory. The UL

Classification Mark for Canada includes: the UL Classification Mark for Canada:  with the word "CLASSIFIED" (as shown), a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) in English, French, or English/French as indicated in the appropriate UL Directory.

Look for the UL Classification Mark on the product

Issued by: *Thorsten Creter*
Thorsten Creter, Project Engineer

UL International Germany GmbH

Any information and documentation provided to you involving UL Mark services are provided on behalf of Underwriters Laboratories Inc.

For questions in Germany, you may call +49 (0) 6102 369 0.

Reviewed by: *Joachim Hochhaus*
Joachim Hochhaus, Project Engineer

UL International Germany GmbH